

EC DECLARATION OF CONFORMITY

MANUFACTURER

Piston Ltd.
Hungary, H-1033, Budapest, Szőlőkert u. 4/b

PRODUCT

Flow meter for Pulmonary Function Test equipment

MODEL NUMBER

PPF-18

CLASSIFICATION

Class IIa
Council Directive 93/42/EEC of MDD, Annex IX

DECLARATION

We herewith declare that the above mentioned products meet the provision of the Council Directive 93/42/EEC for medical devices. All supporting documentation is retained under the premises of the manufacturer

STANDARDS APPLIED

EN 980:2003
EN 1041:1998
ISO 14971:2007
ISO 5356-1:2004

EC CERTIFICATES

ISO 13485:2003	Registration Nr:	SX 60009710 0001
	Report Nr:	02192133 004
CE 0197	Registration Nr:	HD 2111265 01
	Report Nr:	C 2192133 E 01

MAIN ACCESSORIES

PBF-100	Bacterial filter
MPA-30	Mouth piece

BUDAPEST, 12 JAN 2010

Nagy
.....
Kornél NAGY, Mr
Quality Assurance Manager

PISTON Kft
1033 Budapest, Szőlőkert u. 4/b
Adószám: 10465905-2-41